	Application No.	Applicant(s)
Notice of Allowability	10/748,493	LARSEN ET AL.
	Examiner	Art Unit
	Michael P. Alexander	1742
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. This communication is responsive to <u>6 September 2006</u> .		
2. The allowed claim(s) is/are <u>59, 61-66, 68-69 and 73-74</u> .		
3.		
Attachment(s) 1. ☐ Notice of References Cited (PTO-892) 2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948) 3. ☐ Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date 4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material	5. ☐ Notice of Informal P 6. ☐ Interview Summary Paper No./Mail Da 7. ☒ Examiner's Amendr 8. ☒ Examiner's Stateme 9. ☐ Other	(PTO-413), te

Art Unit: 1742

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Lisa Lindquist on 19 September 2006.

The application has been amended as follows: in line 3 of claim 74, change "platinum, chromium, nickel, iron and mixtures thereof" to –platinum, chromium, nickel and iron--.

Allowable Subject Matter

The following is an examiner's statement of reasons for allowance: as indicated in the Office Action of 7 June 2006, the prior art does not specify a method of electropolishing stainless steel, cobalt-chromium alloy or nickel-titanium alloy stents comprising at least one noble metal in an aqueous acidic mixture comprising at least one chelating or complexing agent, said chelating agent comprising at least one sulfur atom and at least one halide in the form of a salt or an acid in combination with the claimed invention.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Application/Control Number: 10/748,493 Page 3

Art Unit: 1742

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael P. Alexander whose telephone number is 571-272-8558. The examiner can normally be reached on M-F 10:00 - 6:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Roy V. King can be reached on 571-272-1244. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MM mpa ROY KING SUPERVISORY PATENT EXAMINER

TECHNOLOGY CENTER 1700